

IS : 2720 (Part XXXIX/Sec 1) - 1977

(Reaffirmed 2002)

Edition 1.2

(1992-10)

Indian Standard

METHODS OF TEST FOR SOILS

**PART XXXIX DIRECT SHEAR TEST FOR SOILS
CONTAINING GRAVEL**

Section I Laboratory Test

(Incorporating Amendment Nos. 1 & 2)

UDC 624.131.377.620.176

© BIS 2004

BUREAU OF INDIAN STANDARDS
MANAK BHAVAN, 9 BAHADUR SHAH ZAFAR MARG
NEW DELHI 110002

Price Group 4

Indian Standard

METHODS OF TEST FOR SOILS

PART XXXIX DIRECT SHEAR TEST FOR SOILS CONTAINING GRAVEL

Section I Laboratory Test

Soil Engineering Sectional Committee, BDC 23

Chairman

PROF DINESH MOHAN

Representing

Central Building Research Institute (CSIR),
Roorkee

Members

ADDITIONAL CHIEF ENGINEER

Public Works Department, Government of Uttar
Pradesh

SHRI D. C. CHATURVEDI (*Alternate*)

ADDITIONAL DIRECTOR RESEARCH
(RDSO)

Railway Board (Ministry of Railways)

DEPUTY DIRECTOR RESEARCH
(RDSO) (*Alternate*)

PROF ALAM SINGH

University of Jodhpur, Jodhpur

LT-COL AVTAR SINGH

Engineer-in-Chief's Branch, Army Headquarters

MAJ R. R. SUDHINDRA (*Alternate*)

DR A. BANERJEE

Cementation Co Ltd, Calcutta

SHRI S. GUPTA (*Alternate*)

CHIEF ENGINEER (D & R)

Irrigation Department, Government of Punjab

DIRECTOR (IPRI) (*Alternate*)

SHRI K. N. DADINA

In personal capacity (*P-820, 'P' New Alipore,
Calcutta 700053*)

SHRI A. G. DASTIDAR

In personal capacity (*5 Hungerford Street, 12/1
Hungerford Court, Calcutta 700017*)

SHRI R. L. DEWAN

Irrigation Research Institute, Khagaul, Patna

DR G. S. DHILLON

Indian Geotechnical Society, New Delhi

SHRI A. H. DIVANJI

Asia Foundations and Construction (P) Ltd,
Bombay

SHRI A. N. JANGLE (*Alternate*)

DR SHASHI K. GULHATI

Indian Institute of Technology, New Delhi

DR G. V. RAO (*Alternate*)

SHRI V. G. HEGDE

National Buildings Organization, New Delhi

SHRI S. H. BALCHANDANI (*Alternate*)

SHRI O. P. MALHOTRA

Public Works Department, Government of Punjab

SHRI J. S. MARYA

Roads Wing, Ministry of Shipping & Transport,
New Delhi

SHRI N. SEN (*Alternate*)

(Continued on page 2)

© BIS 2004

BUREAU OF INDIAN STANDARDS

This publication is protected under the *Indian Copyright Act* (XIV of 1957) and reproduction in whole or in part by any means except with written permission of the publisher shall be deemed to be an infringement of copyright under the said Act.

IS : 2720 (Part XXXIX/Sec 1) - 1977

(Continued from page 1)

<i>Members</i>	<i>Representing</i>
SHRI R. S. MELKOTE SHRI C. SUDHINDRA (<i>Alternate</i>)	Central Water Commission, New Delhi
SHRI T. K. NATARAJAN REPRESENTATIVE RESEARCH OFFICER	Central Road Research Institute (CSIR), New Delhi Hindustan Construction Co Ltd, Bombay Building & Roads Research Laboratory, Chandigarh
SHRI K. R. SAXENA SECRETARY DEPUTY SECRETARY (<i>Alternate</i>)	Engineering Research Laboratories, Hyderabad Central Board of Irrigation & Power, New Delhi
DR SHAMSHER PRAKASH* DR GOPAL RANJAN (<i>Alternate</i>)	University of Roorkee, Roorkee
SHRI H. D. SHARMA SUPERINTENDING ENGINEER EXECUTIVE ENGINEER (<i>Alternate</i>)	Irrigation Research Institute, Roorkee Public Works Department, Government of Tamil Nadu
SHRI B. T. UNWALLA SHRI T. M. MENON (<i>Alternate</i>)	Concrete Association of India, Bombay
SHRI H. C. VERMA SHRI V. K. VASUDEVAN (<i>Alternate</i>)	All India Instruments Manufacturers & Dealers Association, Bombay
SHRI D. AJITHA SIMHA, Director (Civ Engg)	Director General, ISI (<i>Ex-officio Member</i>)

Secretary

SHRI G. RAMAN
Deputy Director (Civ Engg), ISI

Soil Testing Procedures and Equipment Subcommittee, BDC 23 : 3

Convener

PROF ALAM SINGH
University of Jodhpur, Jodhpur

Members

SHRI AMAR SINGH	Central Building Research Institute (CSIR), Roorkee
LT-COL AVTAR SINGH MAJ R. R. SUDHINDRA (<i>Alternate</i>)	Engineer-in-Chief's Branch, Army Headquarters
DEPUTY DIRECTOR RESEARCH (SOIL MECHANICS-I) (RDSO) ASSTT DIRECTOR RESEARCH (SOIL MECHANICS-I) (RDSO) (<i>Alternate</i>)	Railway Board (Ministry of Railways)
SHRI R. L. DEWAN DIRECTOR (I & C) SHRI K. S. PREM (<i>Alternate</i>)	Irrigation Research Institute, Khagaul, Patna Beas Dams Projects, Talwara Township
SHRI H. K. GUHA SHRI N. N. BHATTACHARAYA (<i>Alternate</i>)	Geologist Syndicate Pvt Ltd, Calcutta
DR SHASHI K. GULHATI SHRI R. K. JAIN DR P. K. DE (<i>Alternate</i>)	Indian Institute of Technology, New Delhi United Technical Consultants (P) Ltd, New Delhi
SHRI O. P. MALHOTRA RESEARCH OFFICER (BLDG & ROADS) (<i>Alternate</i>)	Building & Road Research Laboratory, Chandigarh

*He also represents the Institution of Engineers, India.

(Continued on page 7)

Indian Standard

METHODS OF TEST FOR SOILS

PART XXXIX DIRECT SHEAR TEST FOR SOILS CONTAINING GRAVEL

Section I Laboratory Test

0. FOREWORD

0.1 This Indian Standard (Part XXXIX/Sec 1) was adopted by the Indian Standards Institution on 30 September 1977, after the draft finalized by the Soil Engineering Sectional Committee had been approved by the Civil Engineering Division Council.

0.2 With a view to establish uniform procedures for the determination of different characteristics of soils and also for facilitating a comparative study of the results, the Indian Standards Institution is bringing out this Indian Standard methods of test for soils (IS : 2720) which is being published in parts. 38 parts of this standard have been published so far. This part [IS : 2720 (Part XXXIX/Sec 1)-1977] deals with the laboratory determination by direct shear, the shear strength of soils containing gravel with particle size more than 4.75 mm on with disturbed specimen. The test is of two kinds depending upon the state of samples, namely, laboratory test and *in situ* test. The *in situ* test is being covered separately.

0.3 In the formulation of the standard due weightage has been given to international co-ordination among the standards and practices prevailing in different countries in addition to relating it to the practices in the field in this country.

0.4 This edition 1.2 incorporates Amendment No. 1 (September 1987) and Amendment No. 2 (October 1992). Side bar indicates modification of the text as the result of incorporation of the amendments.

0.5 In reporting the result-of a test or analysis made in accordance with this standard, if the final value, observed or calculated, is to be rounded off, it shall be done in accordance with IS : 2-1960*.

1. SCOPE

1.1 This standard (Part XXXIX/Sec 1) covers the method for the laboratory determination by direct shear, of the shear strength of soils containing gravel (with particle size more than 4.75 mm).

NOTE — It is recommended that the 300-mm box shall be used for soils containing gravel up to 30 mm size.

*Rules for rounding off numerical values (*revised*).

IS : 2720 (Part XXXIX/Sec 1) - 1977

1.2 The test shall be carried out at natural moisture content. In case, the deposit is likely to get saturated, the test shall be carried out in the saturated condition.

2. APPARATUS

2.1 The shear box and its assembly shall conform to requirements given in IS : 11593-1986 'Specification for shear box (large) for testing of soils'.

2.2 Loading Device — The major requirements of the loading device are the following :

- a) The vertical stress on the sample shall remain vertical and constant during test. The normal load shall be applied uniformly on the soil specimen in the shear box without eccentricity;
- b) The shear stress or strain shall be applied in the same plane as the dividing plane of the two parts of the shear box;
- c) In case of a stress controlled apparatus, it should be possible to maintain a constant rate of stress increase during the test irrespective of the strain rate; proper arrangement shall be provided to get different rates of stress increase;
- d) In case of strain controlled apparatus, the strain rate shall remain constant irrespective of the stress. Suitable arrangement shall be provided to provide different strain rates; and
- c) No vibrations shall be transmitted to the sample during the test and there shall not be any loss of shear force due to friction between the loading frame and the shear box container assembly.

2.3 Weights (If Necessary) — For providing the normal load through a normal loading device.

2.4 Proving Ring — of suitable capacity fitted with dial gauge accurate to 0.002 mm to measure the shear force.

2.5 Micrometer Dial Gauges — Accurate to 0.01 mm. Two, suitably mounted to measure the horizontal movement and the other two suitably mounted to measure the compression or expansion of the specimen.

2.6 Stop Clock

2.7 Balance — of 50 kg capacity sensitive of 1 kg.

3. PREPARATION OF SPECIMEN

3.1 Specimen may be compacted in layers to the required density by a suitable hammer into the shear box after fixing the two halves of the shear box together by means of fixing screws.

4. PROCEDURE

4.1 The shear box with the soil specimen should be fitted into position as shown in Fig. 1. The required normal load shall be applied. After the required normal load is applied, the shear strain shall be applied. Before the application of shear strain, the upper half of the box should be lifted up slightly to eliminate friction between the parts of the shear box. The shear strain should be applied at a constant rate of 0.2 mm/min on the upper half of the box till the failure of the specimen. The final shear shall be recorded through the calibrated proving ring. At the end of the test, the specimen should be removed from the box and the water content at the shear zone should be determined. The process shall be repeated for the next higher normal load. A minimum of 4 sets of readings shall be taken.

Figures 1 to 4 deleted

5. CALCULATION AND REPORT

5.1 Results of tests shall be recorded suitably. A recommended proforma for recording the result is given in Appendix A.

5.2 The longitudinal displacement at a particular load shall be recorded from the shear displacement dial readings.

5.3 The maximum shear force shall be the peak load from load-displacement curve or where the tangent of flatter portion of later part of the curve leaves in case the curve does not give peak point.

5.4 The maximum shear stress and the corresponding longitudinal displacement (shear displacement) and applied normal stress should be recorded for each test and the result should be presented in the form of a graph in which the applied normal stress is plotted as abscissa and the maximum shear stress is plotted as ordinate. The angle which the resulting straight line makes with horizontal axis and the intercept which the straight line makes with the vertical axis shall be reported as the angle of shearing resistance and cohesion intercept respectively.

NOTE — The normal stress *versus* maximum shear stress relationship may not be straight line in all cases. In such cases the shear parameter shall be obtained by drawing a tangent to the normal stress and maximum shear stress curve at the point of normal stress expected in the field.

APPENDIX A

(Clause 5.1)

PROFORMA FOR RECORDING TEST RESULTS

Project..... Location of sample

Sample No.

Rate of shear strain Proving ring No.

Proving ring constant

Weight of loading frame

Normal load applied

Soil Specimen Measurements

Dimensions Area of specimen.....

Initial wet mass of specimen Volume of specimen.....

Water content

Bulk density

Final wet mass of specimen

Water content at the shear zone.....

Proforma for Recording Shear Stage

i) Thickness of specimenmm ii) Area of cross-section
of specimencm²

iii) Rate of shearingmm/min iv) Normal stress appliedkg/cm²

DATE AND TIME	SHEAR DISPLACEMENT DIAL READINGS	AVERAGE SHEAR DISPLACEMENT	PROVING RING READING	SHEAR FORCE	SHEAR STRESS	VERTICAL DIAL READINGS	AVERAGE VERTICAL DISPLACEMENT
(1)	(2)	(3)	(4)	(5)	(6)	(7)	(8)
	a b					a b	

Plot — Shear stress *versus* shear displacement and find
a) Maximum shear stress at the peak of curve, and
b) Corresponding shear displacement.

Proforma for Recording Summary of Results

TEST NO.	NORMAL STRESS	SHEAR STRESS AT FAILURE	SHEAR DISPLACEMENT AT FAILURE	INITIAL WATER CONTENT	FINAL WATER CONTENT	REMARK
(1)	(2)	(3)	(4)	(5)	(6)	(7)

Plot — Shear stress minus normal stress relationship to obtain
a) Cohesion intercept, and
b) Angle of shearing resistance.

IS : 2720 (Part XXXIX/Sec 1) - 1977

(Continued from page 2)

Members

SHRI R. S. MELKOTE

SHRI C. SUDHINDRA (*Alternate*)

SHRI N. SEN

SHRI P. K. THOMAS (*Alternate*)

SHRI M. M. D. SETH

DR B. L. DHAWAN (*Alternate*)

SHRI P. JAGANNATHA RAO

DR V. V. S. RAO

SHRI H. C. VERMA

PROF T. S. NAGARAJ (*Alternate*)

Representing

Central Water Commission, New Delhi

Ministry of Shipping & Transport (Roads Wing),
New Delhi

Public Works Department, Government of Uttar
Pradesh

Central Road Research Institute (CSIR), New Delhi
In personal capacity (*F-24 Green Park, New Delhi*)

Associated Instruments Manufacturers (I) Pvt Ltd,
New Delhi

Bureau of Indian Standards

BIS is a statutory institution established under the *Bureau of Indian Standards Act, 1986* to promote harmonious development of the activities of standardization, marking and quality certification of goods and attending to connected matters in the country.

Copyright

BIS has the copyright of all its publications. No part of these publications may be reproduced in any form without the prior permission in writing of BIS. This does not preclude the free use, in the course of implementing the standard, of necessary details, such as symbols and sizes, type or grade designations. Enquiries relating to copyright be addressed to the Director (Publications), BIS.

Review of Indian Standards

Amendments are issued to standards as the need arises on the basis of comments. Standards are also reviewed periodically; a standard along with amendments is reaffirmed when such review indicates that no changes are needed; if the review indicates that changes are needed, it is taken up for revision. Users of Indian Standards should ascertain that they are in possession of the latest amendments or edition by referring to the latest issue of 'BIS Catalogue' and 'Standards : Monthly Additions'.

This Indian Standard has been developed by Technical Committee : BDC 23 and amended by CED 23

Amendments Issued Since Publication

Amend No.	Date of Issue
Amd. No. 1	September 1987
Amd. No. 2	October 1992

BUREAU OF INDIAN STANDARDS

Headquarters:

Manak Bhavan, 9 Bahadur Shah Zafar Marg, New Delhi 110002.
Telephones: 323 01 31, 323 33 75, 323 94 02

Telegrams: Manaksanstha
(Common to all offices)

Regional Offices:

	Telephone
Central : Manak Bhavan, 9 Bahadur Shah Zafar Marg NEW DELHI 110002	{ 323 76 17 323 38 41
Eastern : 1/14 C. I. T. Scheme VII M, V. I. P. Road, Kankurgachi KOLKATA 700054	{ 337 84 99, 337 85 61 337 86 26, 337 91 20
Northern : SCO 335-336, Sector 34-A, CHANDIGARH 160022	{ 60 38 43 60 20 25
Southern : C. I. T. Campus, IV Cross Road, CHENNAI 600113	{ 235 02 16, 235 04 42 235 15 19, 235 23 15
Western : Manakalaya, E9 MIDC, Marol, Andheri (East) MUMBAI 400093	{ 832 92 95, 832 78 58 832 78 91, 832 78 92
Branches : AHMEDABAD. BANGALORE. BHOPAL. BHUBANESHWAR. COIMBATORE. FARIDABAD. GHAZIABAD. GUWAHATI. HYDERABAD. JAIPUR. KANPUR. LUCKNOW. NAGPUR. NALAGARH. PATNA. PUNE. RAJKOT. THIRUVANANTHAPURAM. VISHAKHAPATNAM	